

Certificate



Certificate no.

CU 72239128 01

License Holder:

INTELSINT S.r.l.
Via Rivoli 122
10090 Villarbasse TO
Italia

Manufacturing Plant:

INTELSINT S.r.l.
Via Rivoli 122
10090 Villarbasse TO
Italia

Test report no.: USA- IT23WXWW 001

Client Reference: Annarita Serra

Tested to:

UL 61010-1:2012 R7.19
UL 61010-2-101:2019
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
CSA C22.2 No. 61010-2-101:19

Certified Product: Automatic system for preparation of slides

License Fee - Units

for histological analysis (IVD)

Model Designation: CVR (Ref: 920.101)

7

Rated Voltage: AC 115/230V, 50/60Hz

Rated Power: 650W

Protection Class: I

Fuses: 2xT5A 6.3x32mm (500V)

2xT10A 6.3x32mm (500V)

7

Appendix: 1, 1-13

Licensed Test mark:



Date of Issue

(day/mo/yr)

12/04/2023